


Glass Beads



Description: A method used for sterilizing the working ends of endodontic files and reamers or small dental instrument tips by placing them in a container containing glass beads heated to approximately 225°C (437°F) for a defined period of time.

Packing	Product Code
Glass Beads 200 g	5503 / 200